



**QUANTRONIX**  
*Cubing & Weighing Systems*

380 South 200 West  
P.O. Box 929  
Farmington, Utah 84025 USA  
(801) 451-7000

## DECLARATION OF CONFORMITY

---

**Manufacturer Name:** Quantronix Inc.

**Manufacturer Address:** 380 S. 200 W., Farmington, UT 84025 USA

**Declares, that the product**

**Product Model Numbers:** CubiScan 110

**Conforms to the following Product Specifications and Regulations:**

**73/23/EEC**

Electrical equipment for use within specified voltage limits

**89/336/EEC**

Electromagnetic compatibility

The product herewith complies with the requirements of the Low Voltage Directive, the EMC Directive and carries the CE-marking accordingly.

9 June 2011

Signed,

 

Robert Kennington  
Director of Engineering  
Quantronix Inc.

## DECLARATION OF CONFORMITY

References of harmonized standards on which this declaration of conformity is based:

**IEC 61326**

Electrical equipment for measurement, control and laboratory use - EMC requirements Part 1: General requirements

- IEC 61000-4-2 ESD
- IEC 61000-4-3 Radiated Immunity
- IEC 61000-4-4 EFT
- IEC 61000-4-5 Surge
- IEC 61000-4-6 Conducted Immunity
- IEC 61000-4-11 Voltage Dips & Variations

**EN 55011 class A**

Radiated Emissions and Conducted Emissions

**IEC 61000-3-2**

Electromagnetic compatibility (EMC) Part 3: Limits – Section 2: Limits for Current Harmonic Emissions.

**IEC 61000-3-3**

Electromagnetic compatibility (EMC) Part 3: Limits – Section 3: Limits of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current  $\leq 16A$ .

**IEC 60950-1:2005 (2nd Edition) and EN 60950-1: 2006**

Information technology equipment – Safety – Part 1: General requirements

As a result of test reports and their evaluation by accredited laboratories we are in possession of the following certificates for the CubiScan 125:

LABORATORY	REPORT NUMBER
DNB Engineering Inc.	16139A-001