

DECLARATION OF CONFORMITY

Manufacturer Name: Quantronix Inc.

Manufacturer Address: 380 S. 200 W., Farmington, UT 84025 USA

Declares, that the product

Product Model Numbers: CubiScan 125

Conforms to the following Product Specifications and Regulations:

73/23/EEC

Electrical equipment for use within specified voltage limits

89/336/EEC

Electromagnetic compatibility

The product herewith complies with the requirements of the Low Voltage Directive, the EMC Directive and carries the CE-marking accordingly.

1 June 2011

Signed,

 

Robert Kennington
Director of Engineering
Quantronix Inc.

DECLARATION OF CONFORMITY

References of harmonized standards on which this declaration of conformity is based:

IEC 61326

Electrical equipment for measurement, control and laboratory use - EMC requirements Part 1: General requirements

- IEC 61000-4-2 ESD
- IEC 61000-4-3 Radiated Immunity
- IEC 61000-4-4 EFT
- IEC 61000-4-5 Surge
- IEC 61000-4-6 Conducted Immunity
- IEC 61000-4-11 Voltage Dips & Variations

EN 55011 class A

Radiated Emissions and Conducted Emissions

IEC 61000-3-2

Electromagnetic compatibility (EMC) Part 3: Limits – Section 2: Limits for Current Harmonic Emissions.

IEC 61000-3-3

Electromagnetic compatibility (EMC) Part 3: Limits – Section 3: Limits of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current $\leq 16A$.

IEC 60950-1:2005 (2nd Edition) and EN 60950-1: 2006

Information technology equipment – Safety – Part 1: General requirements

As a result of test reports and their evaluation by accredited laboratories we are in possession of the following certificates for the CubiScan 125:

LABORATORY	REPORT NUMBER
DNB Engineering Inc.	16129A-001
DNB Engineering Inc.	16129A-001
Nemko	125491